

Thapar Technology Campus, Bhadson Road  
Patiala-147004, Punjab India  
Phone: +91 175 2393869,70  
Email: [ashwini.aggarwal@thapar.edu](mailto:ashwini.aggarwal@thapar.edu)  
URL: Thapar.edu

**Enquiry No. TIET/CS/AA/21-22/21285**  
**Date : July 21, 2021**

**Sub:** Request for Quotation(s) for the Supply of Spectrum Analyser  
Dear Sir

We shall be grateful if you kindly let us have your lowest **quotations** for the following materials. THE QUOTATIONS SHOULD REACH THE UNDERSIGNED **LATEST BY August 03, 2021** through **courier or e-mail** accompanied by appropriate illustrative literature/catalogues/pamphlets/technical details, samples and specifications as the case may be. **On the quotation envelope/ subject the Enquiry Number & Date should be mentioned on the top of the Envelope/mail subject.**

Sr. No.	Item Name	Qty.
1.	<u>Spectrum Analyser</u> For specifications and qty See attached sheet	01

**The offer sent by you must furnish the following details:**

1. Name, Make & specifications of each item.
2. **Price Breakup itemwise with MRP. (Treat it mandatory)**
3. Educational discount.
4. **Validity of quotation should be at least 60 Days.**
5. **GST %**
6. **Delivery FOR Central Stores TIET, Patiala For imported product CIP New Delhi Airport**
7. Insurance, Freight & other charges if any.
8. Minimum Delivery Period.
9. **Payment terms. Net 30 days against delivery & satisfactory installation at Thapar Institute, Patiala**
10. Guarantee / Warranty Information.
11. **Also please share your Companies Turnover and Market Share along with the offer.**

 Regards,

**Head Commercial**

S.No.	Item(s)	Quantity Required	Specifications
1*	Spectrum Analyser	01	<ul style="list-style-type: none"> <li>• Real Time Spectrum Analysers</li> <li>• Maximum Frequency Range: 4.5 to 6 GHz,</li> <li>• Displayed average noise level (DANL): &lt;-161 dBm,</li> <li>• Phase noise: &lt;-102 dBc/Hz,</li> <li>• Level measurement uncertainty: &lt;1.0 dB)</li> <li>• Handheld or Benchtop</li> </ul>